



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

| APPLICATION NO. | FILING DATE | FIRST NAMED INVENTOR | ATTORNEY DOCKET NO. | CONFIRMATION NO. |
|-----------------|-------------|----------------------|---------------------|------------------|
| 10/069,649 | 09/13/2002 | Aws Nashef | 8830-27 | 3544 |

23973 7590 06/14/2006

DRINKER BIDDLE & REATH
ATTN: INTELLECTUAL PROPERTY GROUP
ONE LOGAN SQUARE
18TH AND CHERRY STREETS
PHILADELPHIA, PA 19103-6996

EXAMINER

GIBSON, ROY DEAN

| ART UNIT | PAPER NUMBER |
|----------|--------------|
|----------|--------------|

3739

DATE MAILED: 06/14/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

| | | | |
|------------------------------|--------------------------------------|--------------------------------------|--|
| Office Action Summary | Application No. 10/069,649 | Applicant(s) NASHEF ET AL. | |
| | Examiner Roy D. Gibson | Art Unit 3739 | |

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 27 March 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 33-57 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 33-37 and 55-57 is/are rejected.
- 7) ☒ Claim(s) 38-54 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Formal Matters

The applicant has submitted a list of patents as an IDS but failed to provide a Form 1449. When this form is provided, the prior art will be formally reviewed and signed by the examiner.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 33-37, 55 and 56 are rejected under 35 U.S.C. 102(b) as being anticipated by Houser (5,865,801).

As to claims 33 and 34, Houser discloses a catheter (12) having a distal end and a wall, the catheter comprising a heat transfer device Figure 8, # 98) located approximately at its distal end wherein the heat transfer device is engaged with a catheter wall and a temperature sensing element (thermocouple with lead wire 102) to measure native blood temperature, wherein the heat transfer device is a flexible film having at least one electrical resistor flow path and includes at least one temperature sensor included on the heat transfer device film to monitor the temperature of the electrical flow paths and thus the temperature of the overall heat transfer device;

Art Unit: 3739

wherein the heat transfer device is a flexible metal film on which the at least one electrical resistor flow path has been formed (col. 4, lines 26-62 and col. 6, lines 40-59).

Further to claims 35-37, Houser discloses:

"Catheter 12 includes thin film heaters attached to the exterior of the balloon adjacent and surrounding the piezoelectric sensors. The thin film heaters are preferably made from a polyamide sheet with a thin resistive coating pattern, e.g., nickel, sputter coated onto the surface. The thin film heaters, one of which is shown at 98 in FIGS. 7 and 8, is connected to a bifilar lead wire 100 and to a thermocouple wire 102 to monitor the temperature. Bifilar lead wire 100 is connected to the opposite ends of a thin film heating element 104 formed on a dielectric sheet 106 of the heater. Thermocouple wire 102 is connected to the thin film heater near the heating element. The leads and thermocouple wire are inserted into the balloon inflation lumen 68 through a hole in the catheter shaft" col. 6, lines 40-49).

Further to claims 55 and 56, Houser discloses the catheter has a diameter of 3-8 F which overlaps the claimed range of 3-5 F and has a single distal lumen (Figure 8 # 80)

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claim 57 is rejected under 35 U.S.C. 103(a) as being unpatentable over Houser.

Although Houser fails to disclose the diameter of the lumen is 0.5-0.7 mm, the examiner maintains that it would have been obvious to a skillful artisan to select this diameter

Art Unit: 3739

range to provide sufficient fluid flow to inflate the balloon particularly and because the catheter diameter is approximately the same as the claimed values of claim 56.

Allowable Subject Matter

Claims 38-54 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).


A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Art Unit: 3739

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Roy D. Gibson whose telephone number is 571-272-4767. The examiner can normally be reached on Tu-Th, 7:30 am-4:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda Dvorak can be reached on 571-272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


Roy D. Gibson
Primary Examiner
Art Unit 3739

June 12, 2006